



DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

SUNGO Europe B.V.
Add: Fascinatio Boulevard 522,
Unit 1.7, 2909VA Capelle aan
den IJssel, The Netherlands
Tel: +31(0)10 3034500 ;
+31(0)2021 11106

Conformity Assessment

Conformity Assessment Procedure

Annex II+III of Regulation (EU) 2017/745

Applicable Standards

EN ISO 14971: 2019
EN ISO 15223-1: 2016
EN ISO 20417: 2021
ISO 10993-1: 2018
EN ISO 10993-5: 2009
EN ISO 10993-10: 2013

Remark

The declaration of conformity is valid in connection with the release technical document CE/MDR-EMSRUN-24.

All the supporting documentation is retained at the premises of the manufacturer.

The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.

Manufacturer

Name: WUXI EMSRUN TECHNOLOGY CO.,LTD

Address: No.100 Fengbin Road,WUXI city,JIANGSU province,China

Product Information

Name: Adhesive bandage

Model: CR-AB01-5

GMDN: 34864

Basic UDI-DI: 697456774CR-AB01QN

Classification: Class I

SRN: CN-MF-000008968

Intended purpose:

Adhesive bandage is used for Rapid hemostasis and trauma nursing.

Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature:  Date: 2021.09.30



Position: GM

Place: WUXI, CHINA